# CRCST Practice Exam (Sample)

**Study Guide** 



Everything you need from our exam experts!

Copyright © 2025 by Examzify - A Kaluba Technologies Inc. product.

#### ALL RIGHTS RESERVED.

No part of this book may be reproduced or transferred in any form or by any means, graphic, electronic, or mechanical, including photocopying, recording, web distribution, taping, or by any information storage retrieval system, without the written permission of the author.

Notice: Examzify makes every reasonable effort to obtain from reliable sources accurate, complete, and timely information about this product.



## **Questions**



- 1. Which of the following package wrapping methods creates a package within a package?
  - A. Single Wrapping
  - **B. Pouch Packaging**
  - C. Sequential
  - D. Paper-Wrapped
- 2. Which of the following is not considered an ergonomic stressor?
  - A. Improper lifting techniques
  - **B.** Job rotations
  - C. Poor posture habits
  - D. Repetitive motions
- 3. Which of the following is not an acceptable method for sealing paper-plastic peels?
  - A. Heat-sealing
  - **B.** Impulse-sealing
  - C. Stapling
  - D. Adhesive tape sealing
- 4. Why is it important to inspect medical instruments after cleaning?
  - A. To ensure they are aesthetically pleasing
  - B. To determine if they need to be replaced
  - C. To confirm they are free of dirt, moisture, and any residual contaminants
  - D. To ensure they are color-coded correctly
- 5. Angstrom is the unit to measure:
  - A. The length of light waves
  - B. The volume of liquids
  - C. The weight of solid objects
  - D. The temperature of gases

- 6. If the surgical team opens a tray and finds the CI missing:
  - A. The tray should be used, and CI found later
  - B. The missing CI should be reported to management only
  - C. The tray may be used if other CIs are present
  - D. The entire tray should be rejected
- 7. Which specialty deals with synovial fluid?
  - A. Ultrasonic cleaners
  - **B.** Orthopedics
  - C. Retractors
  - D. Gelpi
- 8. All these factors impact the effectiveness of sterilization EXCEPT
  - A. Temperature
  - **B.** Time
  - C. Pressure
  - D. Instructions in Manufacturer's IFU
- 9. Which of the following should not be used to remove gross soil on instruments?
  - A. Water
  - B. Alcohol
  - C. Chlorine
  - D. Saline
- 10. Which of the following is the correct sequence for removing Personal Protective Equipment (PPE)?
  - A. Shoe covers, Gloves, Goggles/ face shield, Gown, Mask or respirator
  - B. Gloves, Gown, Shoe covers, Goggles/ face shield, Mask or respirator
  - C. Goggles/ face shield, Mask or respirator, Gloves, Shoe covers, Gown
  - D. Mask or respirator, Gown, Gloves, Shoe covers, Goggles/ face shield

#### **Answers**



- 1. C 2. B 3. C 4. C 5. A 6. D 7. B 8. D 9. D 10. A



## **Explanations**



# 1. Which of the following package wrapping methods creates a package within a package?

- A. Single Wrapping
- **B. Pouch Packaging**
- C. Sequential
- D. Paper-Wrapped

The correct answer is C. Sequential wrapping method creates a package within a package. In this method, items are wrapped individually and then grouped with other items that are also individually wrapped. This creates layers of packaging, with each item enclosed within its own wrapper as well as the outer wrapper formed by grouping the individually wrapped items together. This method helps in maintaining the sterility of each item within the package while also providing an additional layer of protection during storage and transport. The other options are not correct because: A. Single wrapping involves wrapping each item individually without creating a package within a package. B. Pouch packaging involves placing items in a pouch or bag but does not create a package within a package like the sequential wrapping method does.

## 2. Which of the following is not considered an ergonomic stressor?

- A. Improper lifting techniques
- **B. Job rotations**
- C. Poor posture habits
- D. Repetitive motions

Job rotations are not considered an ergonomic stressor because they help reduce ergonomic stress on employees by allowing them to switch between different tasks and positions, thus minimizing the risk of overuse injuries from repetitive motions. In contrast, improper lifting techniques, poor posture habits, and repetitive motions are known ergonomic stressors that can lead to musculoskeletal disorders and injuries in the workplace.

# 3. Which of the following is not an acceptable method for sealing paper-plastic peels?

- A. Heat-sealing
- B. Impulse-sealing
- C. Stapling
- D. Adhesive tape sealing

Stapling is not an acceptable method for sealing paper-plastic peels because it can compromise the sterility of the contents within the peel. Stapling creates punctures in the packaging, which can allow microorganisms to enter and contaminate the items. The other options listed (Heat-sealing, Impulse-sealing, and Adhesive tape sealing) are all acceptable methods for effectively sealing paper-plastic peels without compromising sterility.

- 4. Why is it important to inspect medical instruments after cleaning?
  - A. To ensure they are aesthetically pleasing
  - B. To determine if they need to be replaced
  - C. To confirm they are free of dirt, moisture, and any residual contaminants
  - D. To ensure they are color-coded correctly

Inspecting medical instruments after cleaning is critical to confirm that they are free of dirt, moisture, and any residual contaminants. This step ensures that the instruments are properly cleaned and sanitized, reducing the risk of infection or complications during medical procedures. The presence of any contaminants can jeopardize patient safety and the effectiveness of the instruments, as any remaining particles can interfere with their function or pose health risks to patients. Additionally, this inspection process may identify any potential issues, such as damage or wear, which could impact the instrument's performance. Ensuring that the instruments meet stringent cleanliness standards is a fundamental aspect of infection control in healthcare settings, making this inspection a key part of the overall sterilization and maintenance protocol for medical equipment.

- 5. Angstrom is the unit to measure:
  - A. The length of light waves
  - B. The volume of liquids
  - C. The weight of solid objects
  - D. The temperature of gases

Angstrom is the unit used to measure the length of light waves. It is commonly used in the field of optics and spectroscopy to determine the wavelength of electromagnetic radiation such as visible light. The other options (volume of liquids, weight of solid objects, and temperature of gases) are measured using different units such as liters, grams, and degrees Celsius or Fahrenheit, respectively.

#### 6. If the surgical team opens a tray and finds the CI missing:

- A. The tray should be used, and CI found later
- B. The missing CI should be reported to management only
- C. The tray may be used if other CIs are present
- D. The entire tray should be rejected

If a critical item (CI) is missing from a surgical tray, the entire tray should be rejected. The reason for this is that critical items are essential for the safety and effectiveness of the procedure. Using a tray with a missing critical item could compromise patient safety and the success of the surgical procedure. It is crucial to maintain a high level of quality assurance in the sterile processing department, and part of that involves ensuring that all trays are complete and contain all necessary instruments and supplies. Therefore, the correct course of action when a critical item is missing is to reject the entire tray to prevent any potential risks to the patient and to maintain the integrity of the sterile processing workflow. The other options are incorrect: - Option A is not the correct choice because using a tray with a missing critical item could pose a risk to the patient and compromise the surgical procedure. - Option B is not the best course of action because reporting the missing CI to management only does not address the immediate issue of the incomplete tray. - Option C is not recommended because the absence of a critical item should not be overlooked, even if other CIs are present.

#### 7. Which specialty deals with synovial fluid?

- A. Ultrasonic cleaners
- **B.** Orthopedics
- C. Retractors
- D. Gelpi

Orthopedics is the correct answer because this specialty deals with the musculoskeletal system, which includes joints and synovial fluid. Synovial fluid is a viscous fluid found in the cavities of synovial joints, acting as a lubricant to reduce friction between the articular cartilage of synovial joints during movement. Ultrasonic cleaners, retractors, and Gelpi instruments do not specifically deal with synovial fluid or the musculoskeletal system like orthopedics does.

## 8. All these factors impact the effectiveness of sterilization EXCEPT

- A. Temperature
- **B.** Time
- C. Pressure
- D. Instructions in Manufacturer's IFU

Factors such as temperature, time, and pressure are crucial in determining the effectiveness of sterilization processes. The correct answer, D, 'Instructions in Manufacturer's IFU,' is the exception because these instructions are essential for ensuring proper usage of sterilization equipment according to the manufacturer's recommendations. Following the instructions in the Manufacturer's Instructions for Use (IFU) is necessary for maintaining the effectiveness of the sterilization process and ensuring the safety of the equipment and the items being sterilized. Therefore, while temperature, time, and pressure directly impact the sterilization process, following the manufacturer's instructions is also vital for successful sterilization outcomes.

- 9. Which of the following should not be used to remove gross soil on instruments?
  - A. Water
  - **B.** Alcohol
  - C. Chlorine
  - D. Saline

In the context of cleaning instruments, especially in sterile processing and infection control, it is crucial to use the appropriate materials to remove gross soil effectively. Saline should not be used for this purpose. Saline is primarily a solution of salt in water and is designed for hydration or for certain medical applications, not for cleaning. The use of saline for removing gross soil can potentially leave residues and may not effectively dissolve organic matter, blood, or other debris that can compromise the cleaning process and, ultimately, patient safety. In contrast, water, alcohol, and chlorine can serve specific roles in cleaning and disinfecting processes. Water is effective in initially rinsing off debris, alcohol can help in dissolving certain types of soil, and chlorine is typically used for its disinfecting properties. However, care must be taken with their application to ensure that they do not damage the instruments or leave harmful residues. This understanding of cleaning protocols is vital in the sterile processing field to ensure instruments are properly cleaned and prepared for sterilization, thus maintaining high standards of infection control and patient safety.

- 10. Which of the following is the correct sequence for removing Personal Protective Equipment (PPE)?
  - A. Shoe covers, Gloves, Goggles/ face shield, Gown, Mask or respirator
  - B. Gloves, Gown, Shoe covers, Goggles/ face shield, Mask or respirator
  - C. Goggles/ face shield, Mask or respirator, Gloves, Shoe covers, Gown
  - D. Mask or respirator, Gown, Gloves, Shoe covers, Goggles/ face shield

The correct sequence for removing Personal Protective Equipment (PPE) is crucial in preventing contamination. The correct sequence for removing PPE is to start with removing shoe covers, followed by gloves, goggles/face shield, gown, and finally, the mask or respirator. This sequence ensures that any contaminants on the outer layers of PPE are removed last, reducing the risk of self-contamination. Option B is incorrect because gloves should be removed before the gown to avoid contamination. Option C is incorrect because gloves should be removed before the goggles/face shield. Option D is incorrect because the mask or respirator should be removed last, after the gown.